



MISURE FISCALI

Milan, 25-10-2018


Auditorium TECNIMONT

Master Meter: ALL for ONE, ONE for ALL

All custody / fiscal metering systems have **accuracy** and **repeatability targets** which must be guaranteed during their entire life cycle under **real working conditions**



System Proving



System Proving consists in placing the **duty** meter in **series** with a **meter prover** having sufficiently better measurement accuracy and repeatability than the duty meter

Metering system at factory



high repeatability
high accuracy

Metering system in operation

(bad but possible cases)



low repeatability
low accuracy



high repeatability
low accuracy

- Process Conditions ; fluid, P, T, air etc..
- Measurement technology limits for mechanical meters (wear out, leakages..)

Metering system after Proving



high repeatability
high accuracy


PROVING PROCESS

- 1 **Meter Factor calculation** =
$$\frac{\text{Known Volume (Prover)}}{\text{Unknown Volume (Duty Meter)}}$$
- 2 **Proving Validation criteria**


Reference International Standards for Proving **MPMS Ch.4**

Manual of Petroleum Measurement Standards Chapter 4—Proving Systems


Open Tank Prover



Small Volume Piston Prover



Pipe Prover (Uni or Bi-Di)



Master Meter Prover



Focus on Master Meter Proving

Manual of Petroleum Measurement Standards Chapter 4—Proving Systems

Section 5—Master-Meter Provers

Measurement Coordination Department

FIRST EDITION, OCTOBER 1988



- Third Edition, June 2011
- Note: Latest issue is:
- Fourth Edition, June 2016

- **Master meter proving** is used when proving by the direct method can not be accomplished because of meter characteristics, logistics, time, space, safety, *and cost considerations*.
- The terminology for this type of proving is “**Indirect master meter proving**”; according to MPMS Ch.4.5¶6.5 ,to establish a master meter factor (and this meter could be used for that duty), **a proving against a conventional prover on site or accredited calibration shall be performed** with a repeatability that results in a demonstrated random uncertainty of $\pm 0.029\%$ or better, at a 95% confidence level.
- Our procedure for proving of the master meter to use it then on site normally uses a SVP with criteria of 0.05% repeatability in 5 runs : B+B in the table B.2 examples below:

Combinations	Random Uncertainty (RU1) Value of First Meter	Random Uncertainty (RU2) Value of Second	MMF Range (spread) Defined in 6.2 and 6.3	Combined Uncertainty MMF Uncertainty = $\sqrt{RU_1^2 + RU_2^2} + (0.5 \times \text{MMF Range})$
A+B	0.011 %	0.027 %	0.05 %	0.054 %
A+C	0.011 %	0.073 %	0.10 %	0.124 %
C+B	0.073 %	0.027 %	0.15 %	0.153 %
B+B	0.027 %	0.027 %	0.05 %	0.063 %

Enable a meter to be a Master Meter

Proverdata	Oural crude									
diameter cm	43,19	5,08	1,931E+06	0,000346	0,0000172	94,6814				
ref. Density @15 degr.C	830,3	830,3	830,3	830,3	830,3	830,3	830,3	830,3	830,3	830,3
flowrate tons/hr	160	160	160	160	160	140	140	140	140	140
Omni report	131					132				
prove run	1	2	3	4	5	1	2	3	4	5
pulses	3891,29	3889,991	3891,06	3889,816	3889,305	3887,812	3888,376	3888,348	3887,883	3889,182
prover density kg/M3	819,533	819,504	819,473	819,44	819,402	819,234	819,213	819,193	819,166	819,14
temp. Prover deg.C	29,88	29,92	29,96	30	30,06	30,28	30,31	30,33	30,37	30,4
CTSp	1,000741	1,000742	1,000744	1,000745	1,000747	1,000753	1,000754	1,000755	1,000757	1,000759
CTLp	0,986697	0,986662	0,986625	0,986585	0,986537	0,98634	0,986315	0,98629	0,986258	0,986229
Press. prover Barg	4,02	4,0192	4,016	4,013	4,018	3,98	3,977	3,978	3,978	3,978
CPSp	1,000018	1,000018	1,000018	1,000018	1,000018	1,000018	1,000017	1,000018	1,000018	1,000018
CPLp	1,00034	1,00034	1,00034	1,00339	1,00034	1,000337	1,000337	1,000337	1,000337	1,000337
meter density kg/M3	819,664	819,631	819,602	819,571	819,531	819,35	819,326	819,309	819,287	819,257
temp. Meter deg.C	29,77	29,82	29,86	29,9	29,95	30,18	30,21	30,23	30,26	30,3
CTLm	0,986794	0,986753	0,986753	0,986681	0,986632	0,986429	0,9864	0,986378	0,986352	0,986316
press. Meter Barg	4,761	4,761	4,76	4,757	4,757	4,594	4,591	4,594	4,595	4,595
CPLm	1,000402	1,000402	1,000402	1,000402	1,000402	1,000389	1,000389	1,000389	1,000389	1,000389
correct prover volume l	94,753	94,753	94,754	94,754	94,754	94,754	94,754	94,755	94,755	94,755
prover mass kg	77,6534	77,6508	77,648	77,6449	77,6414	77,626	77,6241	77,6223	77,6199	77,6176
meter mass kg	77,8258	77,7998	77,8212	77,7963	77,7861	77,7562	77,767	77,767	77,7577	77,7836
meterfactor	0,997785	0,998084	0,997774	0,998053	0,99814	0,998325	0,998155	0,99814	0,998228	0,997865
final ref. Density"15 degr. C	833,22	833,22	833,22	833,22	833,22	833,22	833,22	833,22	833,22	833,22
density corr. Factor	1,003516801	1,0035168	1,0035168	1,0035168	1,003516801	1,0035168	1,0035168	1,0035168	1,0035168	1,0035168
vol. corr. Factor provervol. (comp. To Maloney prover)	1,001	1,001	1,001	1,001	1,001	1,001	1,001	1,001	1,001	1,001
meterfactor recalc	1,000293718	1,00059347	1,00028269	1,00056239	1,00064961	1,00083508	1,00066465	1,00064961	1,00073783	1,00037392
error	0,029371772	0,05934695	0,02826901	0,05623916	0,06496103	0,08350754	0,0664648	0,06496103	0,07378316	0,03739189
avg. Meterfactor	1,000476376					1,00065222				
repeatability +/-	0,018346011					0,02305783				
max. deviation	0,036692023					0,04611566				
viscosity kynematic cSt	3,7					3,67				
Reynoldno:	182173					160763				
flowvelocity (measuring tube)	13,16					11,52				
<div> <div>4 inch meter oural crude</div> <div>4 inch meter arabian heavy</div> <div>4 inch meter dalia crude</div> <div>curves 4 inch meter</div> <div>3 inch meter dalia</div> </div>										

∴ repeatability
range = 0.037%

∴ repeatability
range = 0.046%

Master Meter: ONE for ALL, ALL for ONE

Enable a meter to be a Master Meter

N	O	P	Q	R	S	T	U	V	W	X	Y	Z	AA	AB	AC	AD
830,3	830,3	830,3	830,3	830,3		830,3	830,3	830,3	830,3	830,3		830,3	830,3	830,3	830,3	830,3
91	91	91	91	91		60	60	60	60	60		25	25	25	25	25
133						134						135				
1	2	3	4	5		1	2	3	4	5		1	2	3	4	5
3888,388	3888,77	3888,527	3888,354	3888,476		3885,627	3885,874	3886,275	3885,952	3886,065		3885,751	3884,696	3885,271	3884,534	3884,364
819,046	819,025	819,007	818,994	818,966		818,881	818,859	818,821	818,775	818,736		818,666	818,626	818,586	818,548	818,512
30,53	30,56	30,59	30,61	30,64		30,77	30,81	30,85	30,92	30,97		31,07	31,13	31,18	31,23	31,28
1,000767	1,000768	1,00077	1,000771	1,000773		1,000781	1,000783	1,000786	1,000789	1,000792		1,000798	1,000802	1,000806	1,000811	1,000815
0,986109	0,986084	0,986062	0,986046	0,986013		0,985894	0,985866	0,985822	0,985766	0,985719		0,985626	0,985577	0,985529	0,985484	0,985441
4,025	4,029	4,028	4,024	4,027		4,211	4,212	4,215	4,215	4,212		4,305	4,309	4,306	4,304	4,302
1,000018	1,000018	1,000018	1,000018	1,000018		1,000019	1,000019	1,000019	1,000019	1,000019		1,000019	1,000019	1,000019	1,000019	1,000019
1,000342	1,000342	1,000342	1,000342	1,000342		1,000358	1,000358	1,000358	1,000358	1,000358		1,000367	1,000367	1,000367	1,000367	1,000367
819,139	819,108	819,094	819,079	819,056		818,959	818,932	818,896	818,849	818,81		818,743	818,701	818,668	818,627	818,593
30,44	30,48	30,5	30,52	30,55		30,68	30,72	30,77	30,83	30,88		30,98	31,03	31,08	31,13	31,18
0,986197	0,986158	0,986141	0,986123	0,986096		0,985976	0,985941	0,985898	0,985842	0,985795		0,985712	0,985661	0,985621	0,985573	0,985531
4,327	4,332	4,332	4,33	4,331		4,368	4,371	4,374	4,375	4,373		4,39	4,395	4,394	4,394	4,392
1,000367	1,000368	1,000368	1,000367	1,000368		1,000371	1,000371	1,000372	1,000372	1,000372		1,000374	1,000374	1,000374	1,000374	1,000374
94,756	94,756	94,756	94,756	94,756		94,757	94,757	94,758	94,758	94,758		94,759	94,759	94,76	94,76	94,76
77,6093	77,6074	77,6059	77,6022	77,6022		77,5948	77,5929	77,5896	77,5854	77,5819		77,5758	77,5722	77,5688	77,5656	77,5625
77,7678	77,7754	77,7705	77,7695	77,7695		77,7125	77,7175	77,7255	77,719	77,7213		77,715	77,6939	77,7054	77,6907	77,6873
0,997962	0,99784	0,997882	0,997912	0,997848		0,998485	0,998397	0,998251	0,99828	0,998207		0,998208	0,998434	0,998241	0,99839	0,998394
833,22	833,22	833,22	833,22	833,22		833,22	833,22	833,22	833,22	833,22		833,22	833,22	833,22	833,22	833,22
1,0035168	1,0035168	1,0035168	1,0035168	1,0035168		1,0035168	1,0035168	1,0035168	1,0035168	1,0035168		1,0035168	1,0035168	1,0035168	1,0035168	1,0035168
1,001	1,001	1,001	1,001	1,001		1,001	1,001	1,001	1,001	1,001		1,001	1,001	1,001	1,001	1,001
1,00047116	1,00034886	1,00039096	1,00042104	1,00035688		1,00099548	1,00090726	1,00076089	1,00078996	1,00071678		1,00071778	1,00094435	1,00075086	1,00090024	1,00090425
0,04711628	0,0348856	0,03909616	0,0421037	0,03568761		0,09954777	0,09072565	0,07608894	0,07899623	0,07167788		0,07177813	0,09443495	0,07508642	0,09002389	0,09042489
1,00039778						1,00083407						1,0008435				
0,00611534						0,01393495						0,01132841				
0,01223067						0,0278699						0,02265682				
3,64						3,62						3,6				
105382						69880						29286				
7,49						4,94						2,06				

∴ repeatability
range = 0.012%

∴ repeatability
range = 0.028%

∴ repeatability
range = 0.023%

Operating of a Master Meter

Manual of Petroleum Measurement Standards Chapter 4—Proving Systems

Section 8—Operation of Proving Systems

Measurement Coordination
FIRST EDITION, NOVEMBER 1995

- Note: Latest issue is:
- 2nd edition, September 2013

4.8.2.11 PROVER RECALIBRATION FREQUENCY

- c. The maximum interval indicated below has elapsed.
1. Three years for small volume provers and mobile provers.
 2. Five years for permanently installed pipe provers.
 3. Five years for permanently installed tank provers.
 4. Three months for master meter provers.

4.8.2.10 FREQUENCY OF METER PROVING

The frequency required for proving varies from several times a day to twice a year or even longer depending upon the value of the liquid, cost/benefit to prove, meter proving history, meter system stability, and variations of operating systems.

For large volumes or different liquids, a permanently installed prover is normally used. The meters should be proved whenever the flow rate, temperature, pressure, API Gravity (relative density), or viscosity changes significantly. Normally, time or volume is used to determine when the meter should be proved.

The proving frequency for new systems should start at short intervals and be extended to longer intervals as confidence increases in the system.

Coriolis Meter as best selection for Master Meter

- Best traceable **accuracy** and **repeatability** for top class products
- For top class products the traceable **water calibration** on the custody mass meter, is **fully representative of the performance on oil** product in the field then not affected by viscosity and density changes
- **No wear out** problems, no rotative part
- **No need for P or T compensation** for mass proving
- Suitable for mass and volume proving
- Compact and easy to be transported
- **Faster proving** than using displacement provers – and so can be used more often.
- **Easier proving** than using displacement provers – extremely easy to automate and so requires fewer skilled staff in the field.

Operating of a Coriolis Master Meter

- Coriolis meters shall comply with API MPMS Chapter 5.6.
- Prover meter factor for master meter proving of a mass meter using a mass master meter is calculated as below:

$$MF_P = \frac{M_{MM} \times MF_{MM}}{M_M}$$

MF _P	Meter factor calculated from proving	[-]
M _{MM}	Master meter (uncorrected) mass	[tonne]
MF _{MM}	Meter factor of the master meter (at the proving flow rate)	[-]
M _M	Meter on prove (uncorrected) mass	[tonne]

- Process conditions must ensure a single, homogenous fluid
- The meter factor that is determined during proving includes any error the zero value may be contributing. Prior to proving the line meter, but at the operating conditions (pressure and temperature) of the line meter, the master meter observed zero value should be verified.

Case study 1


Region: Far East Asia

Customer Type: Refining

Application: LPG pipeline



Scope of supply :

- N'1 Custody Transfer LPG Pipeline Metering System class
- Z pattern configuration for master meter
- Fully automated proving process
- Fluid: LPG



Master Meter: ONE for ALL, ALL for ONE

Manual of Petroleum Measurement Standards Chapter 4.5 Master Meter Provers




Z pattern / Pay & Check configuration

Approval of Weights & Measures on Site

- Various master-meter verifications at site initiated from control room – fully automated
- Then visit of SIRIM Berhad (wholly-owned company of the Malaysia Government), effectively, the weights & measures agency for Malaysia
- Validation also also from, NGC the end-customer

Name: Mohd Sallehuddin Md Saad N

Date: 29/4/14 D


Signature:  S


(SRC PD)

Name: Zulfazil Bedandun

Date: 29/4/14

Signature: Zulfazil
(NGC)

Name:	SYAHIRUL MANSUR	Yusuf Murtaza
Date:	29/4/14	29/4/14
Signature:		
	(SYML - SYML)	(SYML - SYML)



Shell

Port Dickson, Malaysia

Master Meter Report

29/Apr/2014 11:21

Validation start: 29/Apr/2014 11:16:21

Validation end: 29/Apr/2014 11:20:41

Flow Computer ID: 17-FC-251

Report number: 26

Prove result: Completed

Master factor status: MF implemented

Meter data	
Meter tag nr	17-FCQ-270
Meter manufacturer	E+H
Meter model	PROMASS B4
Meter serial number	H90SE602000
Meter size	1.5
K factor:	1000

Master Meter data	
Meter tag nr	17-FCQ-271
Meter manufacturer	E+H
Meter model	PROMASS B4
Meter serial number	H90SE50200
Meter size	1.5
K factor:	1000

Conditions	
Required successful runs	3
Maximum nr of runs	5
Repeatability limit	1.000 %

Results	
Number of successful runs:	3
Number of performed runs:	3
Run repeatability:	0.322 %

	Mass		Pressure		Temperature		Flow rate	Master factor
	Master tonne	Test tonne	Master bar(g)	Test bar(g)	Master °C	Test °C	tonne/hr	-
Run 1	0.31	0.31	14.55	15.57	40.33	40.47	16	1.000000
Run 2	0.31	0.31	14.61	15.42	40.16	40.32	15	0.996795
Run 3	0.31	0.31	14.37	15.56	40.10	40.25	18	1.000000
Average	0.31	0.31	14.51	15.52	40.19	40.35	16	0.998932

er factor	Value	Deviation %	Limit %	Result
meter factor	0.998932			Passed (Enabled)
ous meter factor	0.997850	0.11	0.25	Passed (Enabled)
priced average	0.998974	0.00	0.25	Passed (Enabled)
meter factor	0.998932	0.00	0.25	Passed (Enabled)

fail	Result
fail	Passed

rk:

by: Mohd Sallehuddin Md Saad

29/4/14

Signature: *[Signature]*

by: Zulkefzal Bekarudin

29/4/14

Signature: *[Signature]*

by: S. S. S. S. S.

29/4/14

Signature: *[Signature]*

by: S. S. S. S. S.

29/4/14

Signature: *[Signature]*

Case study 2

Region: Europe

Customer Type: Market Terminal


Application: Ship Loading

Scope of supply :

- 9 off Custody Transfer MID 0.5 approved systems with Coriolis 12"
- 1 off Coriolis- 12" Master Meter MID approved on portable trailer
- Fluid: Gasoil, Heavy & Mid Bunker Fuel Oil – MGO IFO HFO



Master Meter: ONE for ALL, ALL for ONE



DUTY METER

Calibration Certificate

Certificate No.: 9-4-14-18-0033
Task: 114-26878.02
Certificate date: 2014-10-31
Page: 1 of 1
Appendix: 2

Task specification: Product Verification

Customer equipment:
Make: Endress+Hauser
Type/Class: Promass S4 KXP-FAAS/061ABHN
Serial No.: PCN 80017505-4
Shipped MS id. No.: 323, Seastek Terminal, Antwerp, Belgium
Area of measurement: 310 - 725 t/h
Calibration points: $Q_{min} \leq 0.1 Q_n \leq 0.4 Q_n$; $Q_{max} \leq 0.6 Q_n$
Other: MID EC TEC: DE-09-MD05-PTB 035

Customer: Endress+Hauser
Address: Kägenstrasse 2
4153 Reinach
Switzerland
Alp Cami

Results of the calibration: See appendix 1, meter proving report no 11

Statement of compliance: Full requirements of MID 2004/22/EG MID

Calibrated by: E+H supervised by Force Technology
Supervisor: Thomas Wendebach/Lars Pärmo

Date of calibration: 2014-07-15

Calibrated according to: 7.4.3 7.4.17

Traceability Uncertainty: Appendix 2: SPSE Certificates No 13377, 13

Reference conditions: E+H Master Promass S4K DN350/14", S/N 1

Lars S. Kristensen
MSc Engineering

Lars Pärmo
Technical Manager

FORCE Technology, Div. for Remote Analysis, Measurement og Calibration, Park Alle 345, DK-2605 E
Taastrup, +45 43 26 76 66 Taastrup, +45 43 26 70 14 Email: info@force-tech.com www.force-tech.com

Certificate of Conformity
Measuring Instrument Directive – MID, Module F

Certificate of conformity number: DK-0200-MID05-001-259
Issued by FORCE Certification A/S, Denmark - EC-notified body number 0200

In accordance with Directive 2004/22/EC of the European Parliament and Council of March 31st, 2004 on measuring instruments (MID) with later amendments.

Issued to: Endress+Hauser Instruments International AG
Kägenstrasse 2
4153 Reinach
Switzerland

Reference No.: 114-26878.02

Identification of the measuring system:
Type of instrument: Shipload measuring system
Type designation: Shipload measuring system 323
Serial number: PCN 80017505-4

Conformity assessment is based on the requirements of the regulation and is documented as follows:
EC-type examination certificate: DE-09-MID05-PTB035
Calibration certificate: 9-4-14-18-0033
Other documents: -
Comments: -

Conclusion: The measuring system fulfills the requirements of the regulation.
FORCE Certification identity number 0200 can be affixed to the product in accordance with the procedure in the directive.

Date of issue: Oktober 31, 2014

Lars Pärmo
FORCE Certification, Certification Manager

This certificate of conformity may not be reproduced except in full, without written permission by FORCE Certification A/S.

FORCE Certification A/S - Park Alle 345 - 2605 E Taastrup - Denmark Tel: +45 43 26 71 77 Fax: +45 43 26 71 14 info@forcecertification.com www.forcecertification.com



MASTER METER

CHÂNE D'ETALONNAGE DEBITMETRE LIQUIDE
LIQUID FLOW METER CALIBRATION

ACCREDITATION N° 21340
ACCREDITATION N°

**CERTIFICAT D'ETALONNAGE
CALIBRATION CERTIFICATE**

N° 13377

DELIVRE A: Endress + Hauser
ISSUE FOR:

INSTRUMENT ETALONNE:
CALIBRATED INSTRUMENT

Désignation: Coriolis meter
Designation: Endress et Hauser
Constructeur: Promass S4 KXP
Manufacturer: Promass S4 KXP
Type: 12"
Dimension: 12"
Date de révision: 15/07/2013
Revised date: 15/07/2013
Ce certificat comprend 3 pages et 3 annexes.
This certificate includes 3 pages and 3 appendices.

N° de série: H980CF0200
Serial Number: H980CF0200
N° d'identification: Identification number

Date d'émission: 15/07/2013
Date of issue (dd/mm/yyyy): 15/07/2013

LE RESPONSABLE DU LABORATOIRE
THE HEAD OF THE LABORATORY

Michel FIEU

cofrac
Portes disponibles sur
Rampes disponibles sur
www.cofrac.fr
ETALONNAGE

LA REPRODUCTION DE CE CERTIFICAT S'EST AUTORISEE QUE
Sous LA FORME DE FAC-SIMILE PHOTOGRAPHIQUE INTEGRAL

- Calibration certificate for the duty meter for MID 0.5 application
- Certificate of conformity

- Calibration certificate from accredited laboratory using oil calibration rig with bi directional ball prover
- Yearly verification certificate against our accredited water calibration rig

Case study 3

Region: Middle East

Customer Type: LPG production


Application: LPG pipeline

Scope of supply :

- 1 redundant custody transfer metering system
- 1 installed master meter
- Fully automated process including sampling and proving
- Fluid: LPG



Master Meter: ONE for ALL, ALL for ONE



Case study 4

Region: Africa

Customer Type: Distribution Terminal

Application: truck loading and pipeline metering

Scope of supply :

- 7 – Metering streams for LPG
- 3 – Metering streams for JET FUEL
- 3 – Metering streams for HEAVY GAS OIL
- 23 – Metering streams for GASOLINE
- 23 – Metering streams for GAS OIL
- 1 – Master meter



Case study 5

Region: Turkey

Customer Type: Terminals and Refinery

Application: Various , pipeline and ship loading/unloading

Scope of supply Project :

- Various custody transfer systems for different locations, applications and fluids
- 1 common master meter mounted on truck
- Fluid: Diesel , Gasoil ,LPG, HFO



Test of connection between master meter and metering skid with flexi pipes



Connection test



Grounding plug



Fast plug for signals

Easy to operate as first target in our design



Messages to bring home

- Master metering is a practical and recognized procedure for proving of custody transfer metering systems
- Coriolis meter new generation represent the most indicate and performant technologies for master metering:
 - Full independence from process conditions
 - Highest accuracy both for mass and volume measurement
 - Highest metering stability and repeatability
 - High turndown ratio in order to allow the multiple meter sizes proving with one common device
 - Compact footprint for easy installation on mobile provers

ONE Coriolis master meter for ALL proving needs

(....questions?)